



## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

### Regulations, Acts, Guidances, Notifications, Directives and miscellaneous substances of concern lists

Pall Medical continually monitors the Regulations, Acts, Guidances, Notifications and Directives listed below and in the event of a relevant update occurring, this PSDI will be updated and published to the website. PSDIs will not be updated simply to show reviews have occurred. The substances discussed below are not tested for by Pall Medical.

- Substances included in the REACH Candidate List of Substances of Very High Concern are not known to be present at a concentration > 0.1%
- Substances listed in the REACH restricted substances list (Annex XVII) with relevant restrictions or the REACH authorised substances list (Annex XIV) are not known to be present in the raw materials nor are they intentionally added in the manufacturing process.
- The State of California requires 'clear and reasonable warnings' in respect of amounts of specific chemicals in the consumer products they purchase, in homes or workplaces or that are released into the environment. The aim of the warnings is to protect against chemicals known by the State of California to cause cancer, birth defects or reproductive harm. The list of substances of concern, their form, and/or the exposure level above which notification for each substance is required (or 'safe harbor level') being that published by the State and known as the Proposition-65 list. **These materials can potentially expose users to substances listed below which are known to the State of California to cause cancer and developmental toxicity.**

Substance	CAS#
N,N-Dimethylformamide (DMF)	68-12-2
Vinyl chloride	75-01-4

Pall Medical membranes and filters placed on the market in the State of California are not intended for 'consumer' sale but are for professional use. As the result of use they will be expected to be disposed of as 'hazardous waste' within an appropriate waste stream reflecting the contaminant present as the result of use. These articles are supplied in sealed bags and boxed and any direct contact with the materials of construction of those items is expected to be through 'occupational exposure', which does not require mandatory labelling of articles.

Inhalation or ingestion of filters or sheet or roll membrane articles are considered unlikely scenarios. Also, as gloves are also recommended to be used when handling Pall Medical membranes and filters to maintain cleanliness of the product, skin contact is also considered an unlikely exposure scenario.

- Pall Medical does not specify, intentionally add or have any knowledge of the use of EU RoHS (also known as Directive 2002/95/EC and amendment 2015/863) or any other country RoHS listed materials in the raw materials it purchases, or the intended presence of these materials in any of its products.
- Substances of concern categories listed in the EU Waste Framework are not specified by Pall Medical in the raw materials nor are they intentionally added in the manufacturing processes, but Pall Medical

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

does not test for them. Unused Pall filters are not to the best of our knowledge classified as hazardous waste. Used filter cartridges should be disposed of as clinical waste due to the nature of the contaminants on the filters as a result of use. Therefore, used filters may be classified as hazardous – clinical waste.

- Nanomaterials according to 2011/696/EU and nanomaterial forms of all other substances listed in this PSDI are not known to be present in the raw materials nor are they intentionally added in the manufacturing processes.
- Substances above 0.1 % (weight by weight) which are carcinogenic, mutagenic or toxic (CMR) for reproduction of category 1A or 1B in accordance with the current ATP to Part 3 of Annex VI to Regulation (EC) No. 1272/2008 are not known to be present at a concentration > 0.1%.
- Substances having endocrine-disrupting (ED) properties for which there is scientific evidence of probable serious effects to human health and which are identified either in accordance with the procedure set out in the article 59 of Regulation (EC) No. 1907/2006 or, once a delegated act has been adopted by the Commission pursuant to the first subparagraph of Article 5(3) of Regulation (EU)
- Substances listed in the Stockholm Convention are not known to be present in the raw materials nor are they intentionally added in the manufacturing process.
- The materials of construction and substances employed in making Pall Medical membranes and filters are not specified to have actively anti-microbial, anti-bacterial or other biocidal properties.
- Pall Medical does not specify, intentionally add or have any knowledge of the use of medicinal substances or combinations of substances that are absorbed by or locally dispersed in the human body from its suppliers and does not specify or have any knowledge of the use of such materials in the raw materials it purchases, or the intended presence of these materials in any of its products
- Pall does not specify, intentionally add or have any knowledge of the use of human origin materials including blood or plasma derivatives, in the raw materials it purchases, or the intended presence of these materials in any of its products. However, although Pall Medical takes reasonable precautions to maintain the cleanliness of its products, Pall Medical cannot exclude the possibility of the presence of individual cells originating from human or environmental contact prior to supply of materials to Pall Medical or during the manufacturing and packaging processes, but Pall Medical does not test for them.
- Pall Medical membranes and filters do not knowingly contain materials of direct animal origin i.e. animal parts, tissues, or body fluids or derivatives. They are not known to be present in the raw materials nor are they intentionally added in the manufacturing process.
- Pall Medical membranes and filters do not knowingly contain adjuvants such as mould release agents. They are not known to be present in the raw materials nor are they intentionally added in the moulding or manufacturing processes.
- Nitrosamines and related compounds (Nitrous Acid, secondary, tertiary and quaternary amines, nitrite salts, sodium azide, amine reagents or solvents, nitrate salts that may contain nitrite impurities), are not known to be present in the raw materials nor are they intentionally used in the formulations and manufacturing processes. These processes do not include the use of or the known generation of nitrogen oxides or amines, nor do any of the processes include fermentation steps. There are no specific sourcing controls other than those mentioned above.
- Conflict Minerals as defined below:
  - Tantalum (derived from columbite-tantalite)
  - Tin (derived from cassiterite)
  - Tungsten (derived from wolframite)
  - Gold

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

and their derivatives when originating from the Democratic Republic of Congo, Angola, Burundi, the Central African Republic, Congo, Rwanda, Sudan, Uganda, the United Republic of Tanzania or Zambia, are not known to be present in the raw materials nor are they intentionally added in the manufacturing processes.

- Pall Medical does not specify, intentionally add or have any knowledge of the use of other (non-human or non- animal) non-viable biological substances in the raw materials it purchases, or the intended presence of these materials in any of its products.
- Low levels of vinyl chloride may be present as residuals from membrane raw materials
  
- Substances listed below are not known to be present in the raw materials nor are they intentionally added in the manufacturing processes:
  - 2-MCBT and 2-MBT
  - 4-Nitrotoluene
  - Substances listed in **96/62/EC** and TPCH
  - Alkyl phenols and their ethoxylates
  - Allergens per EU Directive 2003/89 Annex IIa or other known allergens
  - Azo colourants and dyes
  - Asbestos and asbestos fibre
  - Anthracene and its compounds
  - Benzophenones
  - BPA
  - Structural analogues of BPA
  - Bromine or brominated compounds
  - Butylated hydroxy toluene
  - Chlorine
  - Cleaning agents (products are not washed or cleaned during production)
  - Colourants and inks other than those visibly present in or printed on housing or other components.
  - DEHP, DIBP, BBP and DBP
  - Other phthalates of concern
  - Dimethyl fumarate
  - Dinitrobenzenes
  - Dioxins and congeners including polychlorinated and polybrominated dioxins and dibenzodioxins (PCDD and PBDD)
  - Furans and congeners including polychlorinated and polybrominated furans and dibenzofurans (PCDF and PBDF)
  - Flame retardants (halogenated and non-halogenated)
  - Substances derived from Genetically Modified Organisms
  - Hazardous Air Pollutants
  - Heavy metals and their compounds
  - Hydrazine
  - Iso and di-isocyanates

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

- Jatropha derived substances
- MCCPs
- Metals in ink
- Microplastics (any type of plastic fragment used as a raw material that is less than 5 mm in length)
- Natural rubber latex or latex derivatives in the product or packaging
- Organotin compounds
- Ozone depleting substances.
- PBT and vPvB substances
- PFAS
- Non-Phthalate plasticizers
- Pharmaceutical components known to be at risk for melamine contamination.
- Polycyclic aromatic hydrocarbons
- Polystyrene (all types) in product or packaging
- PVC in product
- PVC in packaging
- Radioactive substances
- Recycled materials, post-consumer or other source
- SCCPs
- Sensitizers such as rosin, colophony, R42 and R43
- Silicone containing oils, release agents or sprays.
- Styrene
- Substances derived from micro-organisms (not GMO)
- Thiurams
- Toluene

There are no additional ingredients present which, within the current knowledge of the supplier, are classified and contribute to the classification of the article.

---

### **SECTION 4 - First Aid Measures**

#### **4.1 First aid measures**

Always address any contaminants present on the filter as the result of use.

**Eye Contact:** Eye injury could result from physical impact. Get medical attention immediately.

**Inhalation:** Inhalation is not considered a likely route of exposure for the filter product as supplied by Pall.

**Skin Contact:** Wash with soap and water. If irritation persists, get medical attention.

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

Ingestion: This material is not intended for ingestion and is not expected to present an ingestion hazard in the form and quantities present in a work-place setting. However, if ingestion occurs, seek medical attention.

Protection of first aiders: No action shall be taken involving any personal risk or without suitable training.

### 4.2 Key symptoms and effects

No known significant effects or critical hazards related to the materials of construction of the filter as supplied.

---

## **SECTION 5 - Fire Fighting Measures**

### **5.1 Extinguishing media**

Select an extinguish medium suitable for surrounding / working environment.

For filter set alone use dry chemical, CO<sub>2</sub>, water spray (fog) or foam.

### **5.2 Specific Hazards**

Hazardous thermal decomposition products: CO, CO<sub>2</sub>, acrid smoke, methyl methacrylate

### **5.3 Advice to Fire Fighters**

Special precaution required. Fire-fighters should wear appropriate protective equipment, including self-contained breathing apparatus (SCBA) with a full face-piece operated in positive pressure mode.

---

## **SECTION 6 - Accidental Release Measures**

### **6.1 Personal precautions, protective equipment and emergency procedures**

No special measures are required in respect of the product in the unused condition as supplied.

For used filters always address any contaminants present on the product as the result of use.

### **6.2 Environmental precautions**

For unused product, place in designated waste container appropriate to the materials of construction listed in Section 3 and dispose of in accordance with local regulations via a licenced waste disposal contractor.

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

For used product, using clear-up, containment and appropriate PPE measures related to the product being filtered and the materials of construction detailed in Section 3.

### 6.3 Spillage containment and cleaning up.

Use suitable equipment to collect the used product and place in a designated, labelled waste container.

Care should be taken to consider the nature of any contamination on the product as the result of use and suitable PPE employed for handling medical waste. **See section 13 for disposal requirements.**

---

## SECTION 7 – Handling and Storage

### 7.1 Handling

In the received condition, special protective equipment is not needed during handling and normal use of these products. However, gloves are recommended to prevent contamination of the product and maintain cleanliness. Handling of used product must take into account the nature of potential contaminants.

Put on appropriate personal protective equipment for the working environment (See Section 8). Consult details of product being filtered for specific advice. Avoid activities that can damage the filter/membrane.

Follow good hygiene practices. Eating, drinking and smoking are generally prohibited in areas where this product is handled, stored or processed – exceptions are made on the guidance of local medical advice. Staff must follow standard work-place hygiene before eating, drinking or smoking after using this product. Wear gloves to prevent contamination of the filter/membrane and maintain cleanliness of the unused product.

### 7.2 Storage

The article is supplied dry, without the presence of any preserving fluid. Store in clean, dry conditions suitable for a medical device.

Handle with care to avoid damage.

Do not expose to direct sunlight during storage, or other radiation or direct weather conditions. Store in original shipping bag or boxing.

Ensure careful handling to avoid physical damage. Ensure shipping bag and seals are intact prior to use - do not use if damaged.

Please also consult Pall Medical for further instructions for use information on the product prior to use.

---

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

### **SECTION 8 - Exposure Controls/Personal Protection**

#### **8.1 Control parameters**

Occupational Exposure limits: None required.

Recommended monitoring procedures: None required.

#### **8.2 Exposure controls**

There are no special ventilation requirements for the product as supplied in the new and unused condition.

Hygiene Measures: No special measures required. Good hygiene practice in line with local working environmental requirements and medical guidelines.

Hand protection: Disposable gloves are recommended to ensure product remains clean during installation.

Environmental Exposure Controls: Not normally required for the product itself as supplied.

After the product has been used additional exposure controls care should be taken in line with the nature of any contaminant as a result of its use.

---

### **SECTION 9 - Physical and Chemical Properties**

Appearance: Disposable filter

Physical state: Solid

Colour: Various

Solubility: All components insoluble in water. Acrylic components readily soluble in esters, ketones and chlorinated hydrocarbons

Auto-ignition temperature: Acrylic: 440 °C (830 °F), decomposition begins at

250 °C (482°F)

Versapor membrane: N/A

Sensitive to shock: Mechanical / thermal shock can result in damage.



## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

### **SECTION 10 – Stability and Reactivity**

Reactivity:	The product is stable under the recommended conditions of use and storage.
Chemical Stability:	The product is stable under recommended conditions of use and storage.
Hazardous Polymerisation:	Polymerisation will not occur under recommended conditions of use and storage.
Other hazardous reactions:	Consult details of product being filtered for specific advice. Under normal conditions of storage and use, no hazardous reactions will occur.
Conditions to Avoid:	Avoid hot surfaces or other conditions that soften, swell or adversely affect the product or its materials of construction. Do not allow fluids to freeze on the product.
Incompatible Materials:	Strong Acids, alkalis and oxidising Agents (e.g. Perchloric Acid, nitric acid), alkali metals, strong alkalis and reducing agents, organic acids
Decomposition Products:	Under recommended conditions of use or storage, no hazardous decomposition products will be produced.

---

### **SECTION 11 - Toxicological Information**

The information in this section contains generic advice and guidance in respect of the unused product as supplied. Consult SDS of the product being filtered for specific advice and recommendations.

#### **11.1 Acute Toxicity**

Mutagenicity / Carcinogenicity / Reproductive Toxicity / Teratogenicity: No known concern

Aspiration Hazard: Not applicable.

Potential acute health effects: No known significant effects or critical hazards

#### **11.2 Chronic health effects**

No known significant effects or critical hazards.

Carcinogenicity: No specific test data available, no evidence for hazardous properties

---

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.

## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

### **SECTION 12 - Ecological Information**

Pall Medical products are not expected to degrade in contact with soil or water under ambient conditions.

---

### **SECTION 13 - Disposal Information**

The information in this section contains generic advice and guidance.

#### **Product**

Methods of disposal:

**Hazardous Waste:** To the best of our knowledge, this product if unused is not regarded as hazardous waste as defined by the EU Directive 91/689/EEC and amendments.

**Unused as supplied product:** Disposal should be in-line with national legislation and local regulatory requirements for the materials present. Unused filter/membrane may be used as landfill.

**Used product** should be disposed of as clinical waste due to the nature of the contaminants as a result of use. Therefore, used products may be classified as hazardous – clinical waste.  
Dispose of waste via a licensed waste disposal contractor.

#### **Packaging**

Bagging: Plastic (polyethylene/polyester)

Box: Cardboard

The generation of waste should be avoided or minimised wherever possible. Waste packaging should be recycled where suitable arrangements and facilities exist. Incineration or landfill should only be considered where re-cycling is not feasible.

---

### **SECTION 14 - Transport Information**

The clean and un-used product, supplied in its original packaging, is not classified as dangerous goods under ADR, RID, IMDG or IATA regulations.

---

### **SECTION 15 – Change History**

Rev number	Description of change
1	New document in Pall Medical template.
2	Added 2 part numbers to App 1

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

### **Notice to Reader**

To the best of our knowledge, the information contained herein is accurate. However, Pall Medical assumes no liability whatsoever for the accuracy or completeness of the information contained herein.

Final determination of suitability of any materials is the sole responsibility of the user. All materials may present unknown hazards and should be used with caution. Although certain hazards are described herein, we cannot guarantee that these are the only hazards that exist.

“Pall” is a trademark of the Pall Corporation

® indicates a trademark registered in the USA

© Pall Corporation, 2024

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 2<sup>nd</sup> May 2024

Data Sheet Number: PSDI\_Syringe filters w\_Versapor Family  
Revision: 2

### APPENDIX 1

6094184

6224190

6034199

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.