



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

---

### **SECTION 1 – Product Identification**

This 'Product Safety Data Information' Sheet covers Pall Medical disposable Gas Line filter cartridges, with hydrophobic glass fibre or acrylic copolymer membranes. Unless specified in one of the following sections, all statements refer to all the in-scope part numbers.

Part Number(s): See appendix 1

The filters detailed above are intended for single patient protection in medical gas line filtration applications.

For further information on Pall Medical products, please visit [www.Cytiva.com/pallmedical](http://www.Cytiva.com/pallmedical)

---

### **SECTION 2 - Hazards Identification**

Product definition: Article.

These products are not classified as hazardous according to current versions of UN Recommendations on the Transport of dangerous goods: Model regulations, GB Chemical classification, labelling and packaging (CLP) or European CLP/GHS Regulation 1272/2008 (current ATP)

Signal word: No signal word.

Hazard statements: No known significant effects or critical hazards.

Other hazards: For all listed products except 9004136 and 9004146: Temporary mechanical abrasion (itching) of skin, eyes and respiratory tract may occur on exposure to released fibres or dust. Care should be taken to avoid direct contact or damage to the filter media which would result in free fibre release. Please note persistent glass microfibers used in the manufacture of this product are considered H351 suspected of causing cancer.

Special packaging requirements: None.

---



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

### **SECTION 3 - Materials of Construction**

**3.1** The products detailed in Section 1 are comprised of the following main materials:

Material Name	CAS Number
Modified acrylic copolymer for 8004022, 8004025, 8044023 and ORO family only	Pall proprietary information
Cellulose for BRO2J, 8004022, 8004025, 8044023 and ORO family only	9004-34-6
Ceramic bonded hydrophobic membrane for BRO2J, 8004022, 8004025, 8044023 and ORO family only	Pall proprietary information
Glass Fibre for 6004209, 6004210, 6054210, 6014209, 6154210, 6004605, 6004620 and 6004637 only	Proprietary information
Acrylic copolymer for 9004136 and 9004146 only	Pall proprietary information
Polypropylene for BRO2J, 9004136, 9009146, 6004209, 6004210, 6054210, 6014209, 6154210, 6004605, 6004620 and 6004637 only	9003-07-0
Ethylene Vinyl Acetate (EVA) for ORO1H, ORO1HA and ORO1HJ only.	24937-78-8
Silicone rubber for 6054210 only *	Supplier proprietary information

\* See note on page 5

### **Regulations, Acts, Guidances, Notifications, Directives and miscellaneous substances of concern lists**

Pall Medical continually monitors the Regulations, Acts, Guidances, Notifications and Directives listed below and in the event of a relevant update occurring, this PSDI will be updated and published to the website. PSDIs will not be updated simply to show reviews have occurred. The substances discussed below are not tested for by Pall Medical.

- Substances included in the REACH Candidate List of Substances of Very High Concern are not known to be present at a concentration > 0.1%
- Substances listed in the REACH restricted substances list (Annex XVII) with relevant restrictions or the REACH authorised substances list (Annex XIV) are not known to be present in the raw materials nor are they intentionally added in the manufacturing process.
- The State of California requires 'clear and reasonable warnings' in respect of amounts of specific chemicals in the consumer products they purchase, in homes or workplaces or that are released into the environment. The aim of the warnings is to protect against chemicals known by the State of California to cause cancer, birth defects or reproductive harm. The list of substances of concern, their

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF  
form, and/or the exposure level above which notification for each substance is required (or 'safe harbor level') being that published by the State and known as the Proposition-65 list. **This material can potentially expose users to the substances listed below which are known to the State of California to cause cancer and developmental toxicity.**

Substance	CAS#	Product impacted
Persistent glass microfiber	not listed	BRO2J, 8004022, 8004025, 8044023, ORO family, 6004209, 6004210, 6054210, 6014209, 6154210, 6004605, 6004620 and 6004637
Acetaldehyde	75-07-0	BRO2J, 8004022, 8004025, 8044023, ORO family
Formaldehyde	50-00-0	6004209, 6004210, 6154210, 6004605 and 6004620
1,4 dioxane	123-91-1	BRO2J, 8004022, 8004025, 8044023, ORO family
Methanol	67-56-1	BRO2J, 8004022, 8004025, 8044023, ORO family
Ethyl acrylate	140-88-5	BRO2J, 8004022, 8004025, 8044023, ORO family
Ethylene oxide	75-21-8	BRO2J, 8004022, 8004025, 8044023, ORO family
Propylene oxide	75-56-9	BRO2J, 8004022, 8004025, 8044023, ORO family
Diethanolamine	111-42-2	BRO2J, 8004022, 8004025, 8044023, ORO family
Dichlorohydrin	96-23-1	BRO2J, 8004022, 8004025, 8044023, ORO family
(Epichlorohydrin)	106-89-8	BRO2J, 8004022, 8004025, 8044023, ORO family
3-chloropropane-1,2-diol	96-24-2	BRO2J, 8004022, 8004025, 8044023, ORO family

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

Vinyl Acetate	108-05-4	ORO Family
N,N-Dimethylformamide (DMF)	68-12-2	9004136 and 9004146 only
2-methoxy ethanol	109-86-4	9004136 and 9004146 only
PFOA and its salts	335-67-1	9004136 and 9004146 only
PFOA related compounds	various	9004136 and 9004146 only
Vinyl chloride	75-01-4	9004136 and 9004146 only
Toluene	108-88-3	9004136 and 9004146 only

Pall Medical membranes and filters placed on the market in the State of California are not intended for 'consumer' sale but are for professional use. As the result of use they will be expected to be disposed of as 'hazardous waste' within an appropriate waste stream reflecting the contaminant present as the result of use. These articles are supplied in sealed bags and boxed and any direct contact with the materials of construction of those items is expected to be through 'occupational exposure', which does not require mandatory labelling of articles. Inhalation or ingestion of filters or sheet or roll membrane articles are considered unlikely scenarios. Also, as gloves are also recommended to be used when handling Pall Medical membranes and filters to maintain cleanliness of the product, skin contact is also considered an unlikely exposure scenario.

- Pall Medical does not specify, intentionally add or have any knowledge of the use of EU RoHS (also known as Directive 2002/95/EC and amendment 2015/863) or any other country RoHS listed materials in the raw materials it purchases, or the intended presence of these materials in any of its products.
- Substances of concern categories listed in the EU Waste Framework are not specified by Pall Medical in the raw materials nor are they intentionally added in the manufacturing processes, but Pall Medical does not test for them. Unused Pall filters are not to the best of our knowledge classified as hazardous waste. Used filter cartridges should be disposed of as clinical waste due to the nature of the contaminants on the filters as a result of use. Therefore, used filters may be classified as hazardous – clinical waste.
- Nanomaterials according to 2011/696/EU and nanomaterial forms of all other substances listed in this PSDI are not known to be present in the raw materials nor are they intentionally added in the manufacturing processes.
- Substances above 0.1 % (weight by weight) which are carcinogenic, mutagenic or toxic (CMR) for reproduction of category 1A or 1B in accordance with the current ATP to Part 3 of Annex VI to Regulation (EC) No.1272/2008 are not known to be present at a concentration > 0.1%.
- Substances having endocrine-disrupting (ED) properties for which there is scientific evidence of probable serious effects to human health and which are identified either in accordance with the procedure set out in the article 59 of Regulation (EC) No. 1907/2006 or, once a delegated act has been adopted by the Commission pursuant to the first subparagraph of Article 5(3) of Regulation (EU) No 528/2012 are not known to be present at a concentration > 0.1%,
- In May 2019 the Stockholm Convention moved to add perfluorooctanoic acid (PFOA) and/or PFOA-related compounds to its POPs Annex A, with the intention of restricting the use of PFOA and its

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF related compounds to a limited range of applications within the ratifying countries. For example, European Commission Delegated Regulation 2020/784 and subsequent amendments to Regulation (EU) 2019/1021 regards PFOA, its salts and PFOA-related compounds.

The Versapor RC membrane present in 9004136 and 9004146 only is designed to have hydrophobic and oleophobic properties. The hydrophobic and oleophobic properties of Versapor RC membranes are created by use of a fluorinated additive. According to information received from the additive manufacturer, the additive is not made with PFOA nor is PFOA used in the manufacturing process. The additive comprises a very low percentage of the membrane composition. Pall Medical does not add PFOA in any other part of the formulation or manufacturing process. Pall Medical has conducted independent testing on multiple samples of the additive and a sample of the membrane and PFOA was not detected at a limit of detection/ sensitivity of 10 ppb supporting a conclusion that PFOA is not present in the finished membrane above 25ppb.

- The hydrophobic and oleophobic properties of the membranes are created by use of a fluorinated additive which contains low levels of other PFAS as residuals. Testing of the membrane to determine levels of residual PFAS has been conducted, please refer to the Pall Medical Product Safety Data Information sheet for Versapor RC membrane for further details.
- The materials of construction and substances employed in making Pall Medical membranes and filters are not specified to have actively anti-microbial, anti-bacterial or other biocidal properties.
- Pall Medical does not specify, intentionally add or have any knowledge of the use of medicinal substances or combinations of substances that are absorbed by or locally dispersed in the human body from its suppliers and does not specify or have any knowledge of the use of such materials in the raw materials it purchases, or the intended presence of these materials in any of its products
- Pall does not specify, intentionally add or have any knowledge of the use of human origin materials including blood or plasma derivatives, in the raw materials it purchases, or the intended presence of these materials in any of its products. However, although Pall Medical takes reasonable precautions to maintain the cleanliness of its products, Pall Medical cannot exclude the possibility of the presence of individual cells originating from human or environmental contact prior to supply of materials to Pall Medical or during the manufacturing and packaging processes, but Pall Medical does not test for them.
- Pall Medical membranes and filters do not knowingly contain materials of direct animal origin i.e. animal parts, tissues, or body fluids or derivatives. They are not known to be present in the raw materials nor are they intentionally added in the manufacturing process.
- Pall Medical membranes and filters do not knowingly contain adjuvants such as mould release agents. They are not known to be present in the raw materials nor are they intentionally added in the moulding or manufacturing processes.
- Nitrosamines and related compounds (Nitrous Acid, secondary, tertiary and quaternary amines, nitrite salts, sodium azide, amine reagents or solvents, nitrate salts that may contain nitrite impurities), are not known to be present in the raw materials nor are they intentionally used in the formulations and manufacturing processes. These processes do not include the use of or the known generation of nitrogen oxides or amines, nor do any of the processes include fermentation steps. There are no specific sourcing controls other than those mentioned above.
- Conflict Minerals as defined below:
  - Tantalum (derived from columbite-tantalite)
  - Tin (derived from cassiterite)
  - Tungsten (derived from wolframite)
  - Gold

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

and their derivatives when originating from the Democratic Republic of Congo, Angola, Burundi, the Central African Republic, Congo, Rwanda, Sudan, Uganda, the United Republic of Tanzania or Zambia, are not known to be present in the raw materials nor are they intentionally added in the manufacturing processes.

- Pall Medical does not specify, intentionally add or have any knowledge of the use of other (non-human or non- animal) non-viable biological substances in the raw materials it purchases, or the intended presence of these materials in any of its products.
- Nonylphenol, branched, ethoxylated 68412-54-4 may be present at < 0.1% concentration in BRO2J, 8004022 and ORO family only

Note: Di(2,4-dichlorobenzoyl) peroxide CAS 133-14-2 suspected to be a H317 skin sensitiser may be present in the silicone tubing used on 6054210 at a concentration of < 0.1%

- Substances listed below are not known to be present in the raw materials nor are they intentionally added in the manufacturing processes:
  - 2-MCBT and 2-MBT
  - 4-Nitrotoluene
  - Substances listed in **96/62/EC** and TPCH
  - Alkyl phenols and their ethoxylates not already listed above.
  - Allergens per EU Directive 2003/89 Annex IIa or other known allergens
  - Azo colourants and dyes
  - Asbestos and asbestos fibre
  - Anthracene and its compounds
  - Benzophenones
  - BPA
  - Structural analogues of BPA
  - Bromine or brominated compounds
  - Butylated hydroxy toluene
  - Chlorine
  - Cleaning agents (products are not washed or cleaned during production)
  - Colourants and inks other than those visibly present in or printed on housing or other components.
  - DEHP, DINP, DIBP and BBP
  - Other phthalates of concern.
  - Dimethyl fumarate
  - Dinitrobenzenes
  - Dioxins and cogeners including polychlorinated and polybrominated dioxins and dibenzodioxins (PCDD and PBDD)
  - Furans and cogeners including polychlorinated and polybrominated furans and dibenzofurans (PCDF and PBDF)
  - Flame retardants (halogenated and non-halogenated)
  - Substances derived from Genetically Modified Organisms
  - Hazardous Air Pollutants
  - Heavy metals and their compounds

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

- Hydrazine
- Iso and di-isocyanates
- Jatropha derived substances
- MCCPs
- Metals in ink
- Microplastics (any type of plastic fragment used as a raw material that is less than 5 mm in length)
- Natural rubber latex or latex derivatives in the product or packaging
- Ozone depleting substances.
- PBT and vPvB substances
- Pharmaceutical components known to be at risk for melamine contamination.
- Polycyclic aromatic hydrocarbons
- Polystyrene (all types) in product or packaging
- PVC in product
- PVC in packaging
- Radioactive substances
- Recycled materials, post-consumer or other source
- Rosin, colophony and R42 sensitisers
- SCCPs
- Silicone containing oils, release agents or sprays.
- Styrene
- Substances derived from micro-organisms (not GMO)
- Thiurams

There are no additional ingredients present which, within the current knowledge of the supplier, are classified and contribute to the classification of the article.

---

### **SECTION 4 - First Aid Measures**

#### **4.1 First aid measures**

Always address any contaminants present on the filter as the result of use.

**Eye Contact:** Eye injury could result from physical impact. Get medical attention immediately.

**Inhalation:** Inhalation is not considered a likely route of exposure for the filter product as supplied by Pall.

**Skin Contact:** Wash with soap and water. If irritation persists, get medical attention.

**Ingestion:** This material is not intended for ingestion and is not expected to present an ingestion hazard in the form and quantities present in a work-place setting. However, if ingestion occurs, seek medical attention.

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Protection of first aiders: No action shall be taken involving any personal risk or without suitable training.

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

### 4.2 Key symptoms and effects

No known significant effects or critical hazards related to the materials of construction of the filter as supplied.

---

## **SECTION 5 - Fire Fighting Measures**

### **5.1 Extinguishing media**

Select an extinguish medium suitable for surrounding / working environment.

For filter set alone use dry chemical, CO<sub>2</sub>, water spray (fog) or foam.

### **5.2 Specific Hazards**

Hazardous thermal decomposition products: CO, CO<sub>2</sub>, acrid smoke, hydrogen cyanide, amines, unsaturated hydrocarbons.

### **5.3 Advice to Fire Fighters**

Special precaution required. Fire-fighters should wear appropriate protective equipment, including self-contained breathing apparatus (SCBA) with a full face-piece operated in positive pressure mode.

---

## **SECTION 6 - Accidental Release Measures**

### **6.1 Personal precautions, protective equipment and emergency procedures**

No special measures are required in respect of the product in the unused condition as supplied.

For used filters always address any contaminants present on the product as the result of use.

### **6.2 Environmental precautions**

For unused product, place in designated waste container appropriate to the materials of construction listed in Section 3 and dispose of in accordance with local regulations via a licenced waste disposal contractor.

For used product, using clear-up, containment and appropriate PPE measures related to the product being filtered and the materials of construction detailed in Section 3.

### **6.3 Spillage containment and cleaning up.**

Use suitable equipment to collect the used product and place in a designated, labelled waste container.

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

Care should be taken to consider the nature of any contamination on the product as the result of use and suitable PPE employed for handling medical waste. **See section 13 for disposal requirements.**

---

### **SECTION 7 – Handling and Storage**

#### **7.1 Handling**

In the received condition, special protective equipment is not needed during handling and normal use of these products. However, gloves are recommended to prevent contamination of the product and maintain cleanliness. Handling of used product must take into account the nature of potential contaminants.

Put on appropriate personal protective equipment for the working environment (See Section 8). Consult details of product being filtered for specific advice. Avoid activities that can damage the filter/membrane.

Follow good hygiene practices. Eating, drinking and smoking are generally prohibited in areas where this product is handled, stored or processed – exceptions are made on the guidance of local medical advice. Staff must follow standard work-place hygiene before eating, drinking or smoking after using this product. Wear gloves to prevent contamination of the filter/membrane and maintain cleanliness of the unused product.

#### **7.2 Storage**

The article is supplied dry, without the presence of any preserving fluid. Store in clean, dry conditions suitable for a medical device.

Do not expose to direct sunlight during storage, or other radiation or direct weather conditions.

Store in original shipping bag or boxing.

Ensure careful handling to avoid physical damage.

Ensure shipping bag and seals are intact prior to use - do not use if damaged.

Please also consult Pall for further instructions for use information on the product prior to use.

---

### **SECTION 8 - Exposure Controls/Personal Protection**

#### **8.1 Control parameters**

Occupational Exposure limits: None required.

Recommended monitoring procedures: None required.

#### **8.2 Exposure controls**

There are no special ventilation requirements for the product as supplied in the new and unused condition.

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Hygiene Measures:	Data Sheet Number: PSDI_Gas Line Filter_Family Revision: AF No special measures required. Good hygiene practice in line with local working environmental requirements and medical guidelines.
Hand protection:	Disposable gloves are recommended to ensure product remains clean during installation.
Environmental Exposure Controls:	Not normally required for the product itself as supplied.
After the product has been used additional exposure controls care should be taken in line with the nature of any contaminant as a result of its use.	

---

### **SECTION 9 - Physical and Chemical Properties**

Appearance:	Disposable filter
Physical state:	Solid
Colour:	Clear/White
Solubility:	All components insoluble in water. Acrylic compounds readily soluble in esters, ketones and chlorinated hydrocarbons.
Auto-ignition temperature:	Acrylic: > 400 °C Glass fibre: N/A Cellulose: > 230 °C Polypropylene and EVA: > 300 °C
Sensitive to shock:	Mechanical / thermal shock can result in damage.

---

### **SECTION 10 – Stability and Reactivity**

Reactivity:	The product is stable under the recommended conditions of use and storage.
Chemical Stability:	The product is stable under recommended conditions of use and storage.
Hazardous Polymerisation:	Polymerisation will not occur under recommended conditions of use and storage.
Other hazardous reactions:	Consult details of product being filtered for specific advice. Under normal conditions of storage and use, no hazardous reactions will occur.

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

### Conditions to Avoid:

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF  
Avoid hot surfaces or other conditions that soften, swell or adversely affect the product or its materials of construction. Do not allow fluids to freeze on the product.

Incompatible Materials: Strong Acids, alkalis and oxidising Agents (e.g. Perchloric Acid, nitric acid), alkali metals, strong alkalis and reducing agents, organic acids.

### Decomposition Products:

Under recommended conditions of use or storage, no hazardous decomposition products will be produced.

---

## **SECTION 11 - Toxicological Information**

The information in this section contains generic advice and guidance in respect of the unused product as supplied. Consult SDS of the product being filtered for specific advice and recommendations.

### **11.1 Acute Toxicity**

Mutagenicity / Carcinogenicity / Reproductive Toxicity / Teratogenicity: No known concern for the materials of construction of the filter as supplied (new and unused)

Aspiration Hazard: Not applicable for the unused filter.

Potential acute health effects: No known significant effects or critical hazards

### **11.2 Chronic health effects**

No known significant effects or critical hazards of the unused filter as supplied.

Carcinogenicity: No specific test data available, no evidence for hazardous properties

---

## **SECTION 12 - Ecological Information**

Pall Medical products are not expected to degrade in contact with soil or water under ambient conditions.

---

## **SECTION 13 - Disposal Information**

The information in this section contains generic advice and guidance.

### **Product**

#### Methods of disposal:

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

Hazardous Waste: To the best of our knowledge, this product if unused is not regarded as hazardous waste as defined by the EU Directive 91/689/EEC and amendments.

Unused as supplied product: Disposal should be in-line with national legislation and local regulatory requirements for the materials present. As stated in sections 2 and 3.1, unused filters contain bonded glass fibres which may be released if disposal of unused product occurs without being treated as clinical waste.

Used product should be disposed of as clinical waste due to the nature of the contaminants as a result of use. Therefore, used products may be classified as hazardous – clinical waste.  
Dispose of waste via a licensed waste disposal contractor.

### **Packaging**

Bagging: Polyethylene

Box: Cardboard

The generation of waste should be avoided or minimised wherever possible. Waste packaging should be recycled where suitable arrangements and facilities exist. Incineration or landfill should only be considered where re-cycling is not feasible.

---

### **SECTION 14 - Transport Information**

The clean and un-used product, supplied in its original packaging, is not classified as dangerous goods under ADR, RID, IMDG or IATA regulations.

---

### **SECTION 15 – Change History**

Rev number	Description of change
1	Initial issue replacing Intervene Gas Filter PSDI
2	Addition of 37mm product revising section 2 and 3.1. Correction to CAS# for acetaldehyde and 1,4 dioxane in Proposition-65 table
AC	8004025 and 8044023 added
AD	Added Vinyl Acetate to the California Proposition-65 list following the January 2025 update
AE	Added 6054210 and 6014209 to the composition and Proposition-65 tables in section 3.1. Added an extra row in the composition table for silicone rubber. Added a note on page 5 for skin sensitisers and adjusted the general sensitiser statement at the end of section 3.1. Added 9004136 and 9004146 information to the Proposition-65 table

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



Medical

## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter Family Revision: AF

	(omitted in error), updated the Stockholm convention and PFAS statements and removed toluene from the substance list at the end of section 3.1
AF	Added 6054210, 6014209 to the Polypropylene row in the table in section 3.1 (omission correction)

### **Notice to Reader**

To the best of our knowledge, the information contained herein is accurate. However, Pall Medical assumes no liability whatsoever for the accuracy or completeness of the information contained herein.

Final determination of suitability of any materials is the sole responsibility of the user. All materials may present unknown hazards and should be used with caution. Although certain hazards are described herein, we cannot guarantee that these are the only hazards that exist.

"Pall" is a trademark of the Pall Corporation

® indicates a trademark registered in the USA

© Pall Corporation, 2025

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.



Medical

## PRODUCT SAFETY DATA INFORMATION

Date: 19/11/2025

Data Sheet Number: PSDI\_Gas Line Filter\_Family Revision: AF

### APPENDIX 1

8004022

8004025

8044023

ORO1H

ORO1HA

ORO1HJ

ORO1HNS

BRO2J

9004136

9004146

6004209

6004210

6054210

6014209

6154210

6004605

6004620

6004637

Pall Medical provides the information contained herein in good faith and states that it represents the best information currently available. However, no warranties or representations are expressed or implied and Pall Medical assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Medical be responsible for losses or damages resulting from the use or reliance on this information.