



RESOURCE RPC 1 mL and 3 mL

Reversed phase columns

Instructions for Use

RESOURCE™ RPC 1 mL and 3 mL are prepacked high performance columns for purifying peptides, proteins and other biomolecules by reversed phase chromatography (RPC). The columns are packed with SOURCE™ 15RPC chromatography resin. Use RESOURCE RPC 1 mL for rapid screening. Transfer to RESOURCE RPC 3 mL column for higher resolution and method development on a 10 cm bed height.

Table of contents

1. Description	3
2. Preparation	5
3. Operation.....	7
4. Maintenance	8
5. Ordering information	10

Read these instructions carefully before using the columns.

Intended use

RESOURCE columns are intended for research use only, and shall not be used in any clinical or *in vitro* procedures for diagnostic purposes.

Safety

For use and handling of the product in a safe way, refer to the Safety Data Sheet.

1 Description

SOURCE 15RPC is a polymeric, reversed phase chromatography resin based on rigid, monodisperse ~ 15 µm particles made of polystyrene/divinylbenzene. The polymer surface is underivatized; the hydrophobic interactions between the resin and the sample give a unique selectivity.

SOURCE 15RPC has a high chemical stability. The monodispersity of the particles yields stable beds and gives excellent results at high flow rates. Separations can be run in the pH range 2 to 12, which allows wide flexibility when choosing running conditions. Cleaning can be carried out in the pH range 1 to 14, which makes the use of effective cleaning procedures possible.

The relatively low back pressure versus flow rate makes it possible to use high flow rates. The maximum flow rate is 10 mL/min (1800 cm³/h), while a more typical flow rate range is 1 to 5 mL/min (200 to 900 cm³/h). When you run the column at 5 mL/min in distilled water at 20°C, the backpressure is normally 0.4 MPa (4 bar, 60 psi) for the 1 mL column and about 1.3 MPa (13 bar, 195 psi) for the 3 mL column.

The material of the column hardware is PEEK. The top frit is made of titanium. The bottom filter is made of polyethylene, in the 1 mL column and titanium in the 3 mL column. Table 1 summarizes the characteristics of RESOURCE RPC columns.

SOURCE 15RPC is part of the Cytiva range of BioProcess™ resins. BioProcess chromatography resins are developed and supported for production scale chromatography. BioProcess resins are produced with validated methods and are tested to meet manufacturing requirements. Secure ordering and delivery routines give a reliable supply of resins for production scale. Regulatory Support Files (RSF) are available to assist process validation and submissions to regulatory authorities. BioProcess resins cover all purification steps from capture to polishing.

Table 1. Characteristics of RESOURCE RPC columns.

Resin	SOURCE 15RPC
Matrix	Spherical and monodisperse, porous, rigid, polystyrene/divinyl benzene particles
Mean particle diameter ¹	~ 15 µm
Dynamic binding capacity, Q _{B10} ²	~ 18 mg/mL resin ~ 14 mg Bacitracin/mL resin ~ 45 mg insulin/mL resin
pH stability, operational ³	2 to 12
pH stability, CIP ⁴	1 to 14
Chemical stability	Stable to commonly used aqueous buffers, 1 M HCl, 1 M HCl/90% methanol, 90% acetic acid, 6 M guanidine hydrochloride, 100% n-propanol, 100% ethanol, 100% methanol, 100% acetone, 0.45 M NaOH/40% isopropanol, 1.0 M NaOH ⁵ , 0.1% TFA in water, 0.1% TFA in acetonitrile, 100% isopropanol, 100% tetrahydrofuran
Bed volume	1 mL or 0.3 mL
Column dimension, i.d. x H	
1 mL bed volume	6.4 x 30 mm
3 mL bed volume	6.4 x 100 mm
Maximum operating pressure	4 MPa (40 bar, 580 psi)
Maximum operating flow rate ⁶	10 mL/min
Recommended operating flow rate ⁶	1 to 5 mL/min
Operating temperature	4°C to 40°C
Delivery conditions	20% ethanol
Storage	20% ethanol, 4°C to 30°C

1. Monodisperse size distribution.

2. Dynamic binding capacity at 10% breakthrough by frontal analysis at a mobile phase velocity of 300 cm/h in a HR 10/10 column at 10 cm bed height (2 min residence time) for BSA/bacitracin/insulin in 0.1% TFA in water.

3. pH range where resin can be operated without significant change in function.

4. pH range where resin can be subjected to cleaning- or sanitization-in-place without significant change in function should only be used for cleaning purposes.

5. 1.0 M NaOH should only be used for cleaning purposes.

6. At room temperature using buffers with the same viscosity as water.

2 Preparation

Choosing an eluent system

Since polymer-based resins can be used over a wider pH range and without concerns over mixed mode retention and the need to suppress the ionic interactions of silanol groups, a wider range of eluent protocols can be used.

For samples with unknown properties or known to require acidic conditions

Eluent A: 0.065% TFA in 2% acetonitrile

Eluent B: 0.050% TFA in 80% acetonitrile

For samples known to require basic conditions

Eluent A: 0.125% ammonium solution pH 10 in 2% acetonitrile

Eluent B: 80% acetonitrile in eluent A

Other systems

Eluent A: pH 2.1 0.1% formic acid, 2% acetonitrile

pH 2.0 0.1% acetic acid, 2% acetonitrile

pH 2.0 0.1% TFA, 2% acetonitrile

pH 4.5 10 mM sodium acetate, 2% acetonitrile

pH 7.0 10 mM potassium phosphate, 2% acetonitrile

pH 9.0 10 mM Tris-HCl, 2% acetonitrile in buffer

pH 12 10 mM NaOH, 2% acetonitrile in buffer

Eluent B: 70% acetonitrile in eluent A

Preparing eluent and sample

We strongly recommend that you prepare solvents and samples carefully to protect the column.

Preparing eluent and sample:

- 1** Filter eluents which have had solids added, using a 0.22 μm filter. This prevents particles from clogging the column.
- 2** Measure volumes of organic solvent and aqueous solutions separately and then mix (this eliminates volume variations which occur when mixing organic and aqueous phases directly).
- 3** Degas the solutions in a sonication bath (< 15 min), under vacuum with magnetic stirring (< 5 min) or by purging with helium (< 5 min). This prevents bubble formation during elution. Be careful to keep the degassing time as low as possible in order to prevent evaporation of the organic solvent.
- 4** Add volatile ion pairing agents.
- 5** Dissolve the sample in eluent A.
- 6** Centrifuge samples at 10 000 g for 10 min or filter through a 0.22 or 0.45 μm sterile filter. Use a solvent-resistant filter if there is an organic modifier in eluent A. Apply to the column as soon as possible to avoid any side reactions such as oxidation.

3 Operation

Connecting the column

Before connecting the column, wash the liquid pathway of the system with the eluents you plan to use. This makes sure that none of the previous solutions remain in the tubings.

Always use a flow restrictor (compatible with an appropriate pressure range) connected after the detector of a chromatography system to prevent the accumulation of air in the detector.

Note: *RPC is not recommended for protein purifications if recovery of activity and return to a correct tertiary structure are required, since many proteins are denatured in the presence of organic solvents.*

Equilibrating the column

The column is delivered in 20% ethanol. When you equilibrate the column for first time use, after long term storage, or when changing eluents, proceed according to steps 1 to 3 below.

- 1 Wash the column with approximately 3 column volumes of eluent B at a low to moderate flow rate.
- 2 Run a 2 to 3 column volume linear gradient from 100% eluent B to 100% eluent A at the same flow rate as in step 1.
- 3 Equilibrate the column with 10 column volumes of eluent A. Continue equilibration until all monitor signals are stable.

Separation by gradient elution

Flow: 1 to 5 mL/min

Collect fractions throughout the separation.

- 1 Equilibrate the column with at least 10 column volumes of eluent A until the UV signal is stable.
- 2 Dissolve the sample in a small volume of eluent A. Filter or centrifuge to remove particulate matter if necessary. Apply to the column.

- 3 When the UV signal is stable, that is, when all unbound material has washed through the column, elute using a gradient of 10 to 20 column volumes from 0% to 100% eluent B.
- 4 Wash the column with at least 5 column volumes of 100% eluent B (or until UV signal is stable) to elute any remaining material.
- 5 Wash with a gradient of 2 to 3 column volumes from 0% to 100% eluent B.
- 6 Re-equilibrate with 10 column volumes of eluent A or until UV signal is stable.

4 Maintenance

Storage

When you have finished using the column, wash it with at least 10 column volumes of distilled water, equilibrate with at least 10 column volumes of 20% ethanol. Store the column at 4°C to 30°C. Make sure that the column is sealed well to avoid drying out. Do not freeze.

Cleaning

Correct preparation of samples and eluents, including filtration, the removal of any particulate matter and a final wash step in 100% eluent B, should keep most columns in good condition. However, reduced performance, reduced flow, increasing back pressure or complete blockage are all indications that the resin needs to be cleaned using more stringent procedures in order to remove tightly bound, precipitated or denatured substances.

It is not recommended to reverse the direction of flow due to the column design. The number of column volumes and contact time required for each cleaning step might vary according to the degree of contamination.

Contact time, organic solvent and pH are significant parameters for successful cleaning, and different protocols might have to be developed and used in combination according to the nature of the contaminants.

Examples of cleaning protocols are as follows:

Eluent A: 0.1% TFA

Eluent B: 0.1% TFA in 80% acetonitrile

Flow: 1.0 mL/min

- 1 Equilibrate the column with at least 10 column volumes of eluent A until the UV signal is stable.
- 2 Wash using a gradient of 20 to 30 column volumes from 0% to 100% eluent B.
- 3 Wash the column with at least 10 column volumes of 100% eluent B.
- 4 Wash using a gradient of 20 to 30 column volumes from 0% to 100% eluent B.
- 5 Wash the column with at least 10 column volumes of eluent A.
- 6 Equilibrate the column in at least 10 column volumes in the eluent A that will be used for the separation if different the eluent used in step 5. Transfer between the two eluents must be performed using a 2 to 3 column volume gradient if the two eluents are significantly different.

Change to 0.1%TFA in 2-propanol for eluent B if column performance is not restored. Note that 2-propanol will increase backpressure, and flow rates might need to be reduced.

For removal of contaminants known to be acid- or alkali-soluble the following eluents can be used, according to the same procedure as outlined above:

Removal of acid-soluble contaminants

Eluent A: 90% acetic acid

Eluent B: 80% acetonitrile or 50% 2-propanol

Removal of alkali-soluble contaminants

Eluent A: 0.5 M NaOH

Eluent B: 50% acetonitrile or 50% 2-propanol

If neither of the protocols for acid- or alkali-soluble contaminants is successful, wash the column in 5 to 10 column volumes of 6 M guanidine hydrochloride.

5 Ordering information

Product	Quantity	Product code
RESOURCE RPC 1 mL	1	17118101
RESOURCE RPC 3 mL	1	17118201

Accessories	Quantity	Product code
Union M6 female/1/16" male (for connection to FPLCTM systems)	5	18385801
Fingertight connector 1/16" (for connection to ÄKTA™ design systems)	10	18111255

Related Products	Quantity	Product code
SOURCE 15RPC	10 mL	17072720
Hydrophobic Interaction & Reversed Phase Chromatography, Principles and Methods		11001269

Page intentionally left blank



cytiva.com

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate.

ÄKTA, BioProcess, RESOURCE and SOURCE are trademarks of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva.

All other third party trademarks are the property of their respective owners.

© 2020 Cytiva

All goods and services are sold subject to the terms and conditions of sale of the supplying company operating within the Cytiva business. A copy of those terms and conditions is available on request. Contact your local Cytiva representative for the most current information.

For local office contact information, visit cytiva.com/contact.

71717400 AL 06/2020