

Efficiency test of ReadyToProcess columns

Intellectual Property Notice: The Biopharma business of GE Healthcare was acquired by Danaher on 31 March 2020 and now operates under the Cytiva[™] brand. Certain collateral materials (such as application notes, scientific posters, and white papers) were created prior to the Danaher acquisition and contain various GE owned trademarks and font designs. In order to maintain the familiarity of those materials for long-serving customers and to preserve the integrity of those scientific documents, those GE owned trademarks and font designs remain in place, it being specifically acknowledged by Danaher and the Cytiva business that GE owns such GE trademarks and font designs.

cytiva.com

GE and the GE Monogram are trademarks of General Electric Company.

Other trademarks listed as being owned by General Electric Company contained in materials that pre-date the Danaher acquisition and relate to products within Cytiva's portfolio are now trademarks of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva.

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate. All other third-party trademarks are the property of their respective owners. © 2020 Cytiva

All goods and services are sold subject to the terms and conditions of sale of the supplying company operating within the Cytiva business. A copy of those terms and conditions is available on request. Contact your local Cytiva representative for the most current information.

For local office contact information, visit cytiva.com/contact

CY13256-08May20-AN

Efficiency test of ReadyToProcess columns

Key words: Efficiency test, Theoretical plate, Asymmetry factor, Process-scale chromatography

Abstract

ReadyToProcess™ columns are prepacked, pre-qualified, and intended for immediate use. Every individual column is qualified by efficiency testing, which comprises analysis of theoretical plates per m packed bed (N/m) and asymmetry factor (As). Acceptance limits have been established for efficiency testing at 100 cm/h, which is a higher liquid velocity than that usually recommended in order to obtain the maximal calculated plate number/m packed bed. This application note compares the results obtained for different columns and sizes at 30 cm/h and 100 cm/h using a BioProcess™ system with a peristaltic pump.

Although the values obtained at 100 cm/h are lower than values obtained at 30 cm/h, the difference is predictable. The values obtained at 100 cm/h are directly comparable with those given in the certificates accompanying ReadyToProcess columns, and better reflect the efficiency during operation.

Introduction

ReadyToProcess columns are prepacked, pre-qualified, and pre-sanitized process chromatography columns suited for purification of biopharmaceuticals (e.g., proteins, vaccines, plasmids, and viruses) for clinical phase I and II studies. ReadyToProcess columns are available with several media at different volumes (2.5, 10, and 20 l).

The quality of packed columns is frequently expressed in terms of number of theoretical plates/m packed bed (N/m) and asymmetry factor (A_c) . These values are calculated as follows:



Fig 1. Typical test chromatogram showing asymmetry factor $({\rm A_s})$ value calculations.

Number of theoretical plates, N

 $N = 5.54 \times (V_R/w_h)^2$, assuming a Gaussian peak where

N = number of theoretical plates

 V_{p} = peak retention (elution) volume or time

 $w^{}_{\rm h}$ = peak width at half height expressed in the same units as $V^{}_{\rm R}$

Asymmetry factor $A_s = b/a$

where

a is partial peak width, measured at 10% of the peak height for the leading part of the peak (Fig 1) and

b is partial peak width, measured at 10% of the peak height for the tailing part of the peak.





Experimentally, both values are easily determined in an efficiency test, by applying a tracer (sample) such as acetone. The calculated theoretical plates/m varies depending on the test conditions and should therefore be regarded as a reference value. To achieve comparable results, conditions and equipment should be kept as constant as possible since changes in buffers, sample volume, liquid velocities, liquid pathway, temperature, etc. will influence the result. For more information of principles for efficiency testing in general, see Reference 1.

To ensure a well-packed column, each individual ReadvToProcess column is aualified by analysis of theoretical plates/m packed bed and asymmetry factor as a part of the production procedure. Acceptance limits have been established for efficiency testing at 100 cm/h and the analysis results are given in a Certificate of Analysis accompanying each column. This is a higher liquid velocity than that normally recommended for optimal determination of theoretical plates/m packed bed (10–50 cm/h), and is a balance between better reflecting the running conditions used during separation, and the possibility to obtain reliable values. The values will differ between media due to the effect of the particle size; larger particles give a lower plate number.

This application note describes a reliable and reproducible efficiency test that can be employed by the user of ReadyToProcess columns.

Material and methods

The column to be tested was attached to a BioProcess system with a peristaltic pump and UNICORN™ software 5.11. The system configuration is described in Figure 2 and the method used is shown in Figure 3. The columns are designed for upwards flow. Three repeated runs were made on each column. Theoretical plates/m packed bed and asymmetry factor were calculated by using the peak integration function in the software evaluation mode.

Results and discussion

Theoretical plates/m packed bed and asymmetry factor values have been determined for ReadyToProcess columns packed with different separation media and with various volumes. The results obtained at a liquid velocity of 100 cm/h



Fig 2. Schematic drawing of the system configuration.

Running conditions

Sample (Inlet Sample):	2% acetone (v/v) in distilled water
Eluent (Inlet 1):	Distilled water
Liquid velocity:	100 cm/h (upflow)
Column:	Equilibrated with distilled water

. . .

Running method					
Step 1. Priming the sys	tem with sample				
Inlet:	Sample				
Outlet:	Waste				
Air trap:	Bypass				
Column:	Bypass				
Duration:	Fill the system with sample (check UV signal)				
Step 2. Apply sample, 1	Step 2. Apply sample, 1% of column volume, to column				
Inlet:	Sample				
Outlet:	Waste				
Air trap:	Bypass				
Column:	Inline				
Duration:	1% of CV				
Method base:	Time				
Step 3. Wash out sample, fill the system with eluent					
Inlet:	1				
Outlet:	Waste				
Air trap:	Bypass				
Column:	Bypass				
Duration:	Fill the system with eluent				
	(check UV signal)				
Step 4. Elute sample					
Inlet:	1				
Outlet:	Waste				
Air trap:	Inline				
Column:	Inline				
Duration:	1.3 CV				

Fig 3. Running conditions recommended for efficiency tests of ReadyToProcess columns.



Fig 4. Average theoretical plates/m packed bed and asymmetry factor calculated for repeated runs performed on 10 ReadyToProcess columns at a liquid velocity of 30 and 100 cm/h respectively. The bars represent different column volumes, media, and natural variation in packing quality within specification. The standard deviation of asymmetry factor was lower than 0.06.

Table 1. Theoretical plates/m packed bed and asymmetry factor calculated for RTP Capto S $20\,$

Table 2. Ratio of theoretical plates/m packed bed obtained at 100 cm/h and 30 cm/h, respectively

Liquid velocity (cm/h)	Theoretical plates/m packed bed (N/m)	Asymmetry factor (A _s)
30	5432	1.13
100	3371	1.08

have been compared with values obtained from equivalent experiments run at 30 cm/h, which is a normal liquid velocity for efficiency tests (Fig 4, Table 1).

When the test was performed at a liquid velocity of 100 cm/h, an increase in band broadening was seen when compared to the same test performed at 30 cm/h (Fig 5). The difference depends on mass transfer effects between the mobile phase and the stationary phase. The asymmetry factor is often improved at higher liquid velocities. This is probably due to a better liquid distribution over the packed bed and a symmetric broadening of the peak. It was found that the asymmetry factor is larger for small columns than for large ones. The system contributes with some asymmetry and the total asymmetry is proportionally larger for small columns. A larger sample volume, 2.5% of the column volume, will decrease the asymmetry factor, and is recommended for efficiency testing.

As expected, the plate numbers achieved at 100 cm/h differ from the plate numbers obtained at 30 cm/h. The values obtained at 100 cm/h were between 50% and 65% of values obtained at the lower liquid velocity (Fig 4, Table 2). This result is consistent for many separation media and column sizes.

Column: RTP Capto™ S 20, bed height 20 cm Equilibration solvent: Distilled water Sample 2% acetone in distilled water Sample load 200 ml Distilled water Eluent: Liquid velocity: 30 or 100 cm/h System: BioProcess system with peristaltic pump mAU 100 80 60 40 20 0 10000 5000 15000 20000 ml

Fig 5. Sample peak with a liquid velocity of 30 cm/h (blue curve) and 100 cm/h (red curve).

Column	Ratio 100:30 cm/h flow rates (%)
RTP MabSelect SuRe™ 2.5	65
RTP MabSelect SuRe 20	65
RTP Phenyl Sepharose™ 6 FF (low sub) 2.5	52
RTP Phenyl Sepharose 6 FF (low sub) 20	50
RTP Capto Q 2.5	60
RTP Capto Q 20	56
RTP Capto S 2.5	58
RTP Capto S 20	59
RTP Capto adhere 10	63
RTP Capto adhere 20	60

Conclusions

This application note compares the results from efficiency testing of different ReadyToProcess columns and sizes obtained at 30 cm/h and 100 cm/h, respectively, using a BioProcess™ system with a peristaltic pump. The values obtained for theoretical plates/m at 100 cm/h are lower than the corresponding values obtained at 30 cm/h, but the variation is predictable. Generally, 100 cm/h gives better asymmetry (closer to 1.0). The data for 100 cm/h are representative for the values obtained in the column certificates for ReadyToProcess columns. Since the system contribution to the asymmetry factor value is diminished at higher sample volumes, production testing is performed with a larger sample volume for similar measurements.

Reference

 Sofer, G. and Hagel, L. Eds., Handbook of process chromatography: A guide to optimization, scale-up and validation. Academic Press, San Diego (1997).

Ordering information

Product	Column size	Code No.
RTP Capto Q 2.5	2.5	28-9017-23
RTP Capto Q 10	10	28-9017-24
RTP Capto Q 20	20	28-9017-25
RTP Capto S 2.5	2.5	28-9017-29
RTP Capto S 10	10	28-9017-30
RTP Capto S 20	20	28-9017-31
RTP Capto adhere 2.5	2.5	28-9017-14
RTP Capto adhere 10	10	28-9017-15
RTP Capto adhere 20	20	28-9017-16
RTP MabSelect SuRe 2.5	2.5	28-9017-17
RTP MabSelect SuRe 10	10	28-9017-18
RTP MabSelect SuRe 20	20	28-9017-19
RTP Phenyl Sepharose 6 FF (low sub) 2.	5 2.5	28-9017-35
RTP Phenyl Sepharose 6 FF (low sub) 10	D 10 I	28-9017-36
RTP Phenyl Sepharose 6 FF (low sub) 20	201	28-9017-37

For contact information for your local office, please visit, www.gelifesciences.com/contact

www.gelifesciences.com/bioprocess

GE Healthcare Bio-Sciences AB Björkgatan 30 751 84 Uppsala Sweden



imagination at work

GE, imagination at work, and GE monogram are trademarks of General Electric Company.

BioProcess, Capto, Drop design, MabSelect SuRe, ReadyToProcess, and Sepharose are trademarks of GE Healthcare companies.

Separating viral particles with Capto Q products may require a license under United States patent number 6,537,793 B2 and equivalent patents and patent applications in other countries owned by Centelion SAS. Such a license is not included with the purchase of Capto Q but is included with the purchase of Capto ViralQ products.

All third party trademarks are the property of their respective owners. © 2007 General Electric Company—All rights reserved First published Sept. 2007.

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is avoilable on request. Contact your local GE Healthcare representative for the most current information.

GE Healthcare UK Ltd Amersham Place, Little Chalfont, Buckinghamshire, HP7 9NA, UK GE Healthcare Bio-Sciences Corp

GE Healthcare Bio-Sciences Corp. 800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA GE Healthcare Europe GmbH Munzinger Strasse 5, D-79111 Freiburg, Germany

Munzinger Strasse 5, D-79111 Freiburg, Germany GE Healthcare Bio-Sciences KK Sanken Bldg., 3-25-1, Hyakunincho, Shinjuku-ku Tokyo, 169-0073 Japan