

HiLoad 16/600 and 26/600 Superdex 30 prep grade

HiLoad 16/600 and 26/600 Superdex 75 prep grade

HiLoad 16/600 and 26/600 Superdex 200 prep grade

Instruction for Use

Introduction

HiLoad™ 16/600 and 26/600 Superdex™ 30 prep grade, HiLoad 16/600 and 26/600 Superdex 75 prep grade, and HiLoad 16/600 and 26/600 Superdex 200 prep grade (pg) are prepacked XK columns designed for preparative size exclusion chromatography.

Superdex prep grade is a composite matrix of dextran and cross-linked agarose. The steep selectivity of dextran and the high chemical and physical stability of agarose enable high resolution separations. Steep selectivity curves give unmatched resolution for biomolecules in the molecular weight range up to ~ 10 000 for Superdex 30 pg, ~ 3000 to 70 000 for Superdex 75 pg, and ~ 10 000 to 600 000 for Superdex 200 pg (Fig. 2, on page 1).

The chromatography resins combines high mechanical strength with high hydrophilicity, allowing high flow rates and minimal non-specific interactions.

Table 1. Contents of the delivery box

| Component | No. supplied |
|-----------------------|--------------|
| Transport device | 1 |
| 1/16" male connectors | 2 |
| Stop plug | 1 |
| HiLoad column | 1 |
| Instructions | 1 |




Fig 1. Package includes HiLoad column, transport device, two connectors, two stop plugs and instructions.






Fig 2. Selectivity curves from Superdex 30 pg, Superdex 75 pg and Superdex 200 pg.

Table 2. HiLoad column characteristics

| | |
|---|--|
| Matrix | Cross-linked agarose, spherical |
| Particle size, d50V ¹ | ~34 µm |
| Fractionation range (M _r) | <10 000 (Superdex 30 pg) |
| Globular proteins | ~3 × 10 ³ to 7 × 10 ⁴ (Superdex 75 pg) |
| Dextrans | ~1 × 10 ⁴ to 6 × 10 ⁵ (Superdex 200 pg) ~5 × 10 ² to 3 × 10 ⁴ (Superdex 75 pg) ~1 × 10 ³ to 1 × 10 ⁵ (Superdex 200 pg) |
| Column volume ² | 120 to 124 mL (16/600) 319 to 330 mL (26/600) |
| Sample volume ³ | Up to 5 mL (16/600) Up to 13 mL (26/600) |
| Theoretical plates | >13 000 m ⁻¹ |
| Maximum operating pressure, Δp | 0.3 MPa (3 bar, 42 psi) |
| Column hardware pressure limit ⁴ | 0.5 MPa (5 bar, 73 psi) |
| pH stability, operational ⁵ | 3 to 12 |
| pH stability, CIP ⁶ | 1 to 14 |
| Storage: | |
| • Superdex 30 | • 0.2 M sodium acetate, 20% ethanol at 4°C to 30°C |
| • Superdex 75 | • 0.2 M sodium acetate, 20% ethanol at 4°C to 30°C |
| • Superdex 200 | • 20% ethanol at 4°C to 30°C |

¹ Median particle size of the cumulative volume distribution.² The surface of the resin is not directly visible at the bottom piece. Therefore, when calculating the total column volume, calculate the height of the resin from the lowest part of the bottom piece to the surface of the resin/adapter. For HiLoad 16/600 deduct 30 mm, and for HiLoad 26/600 deduct 36 mm.³ Optimal sample volume depends on the complexity of the sample and the flow rate. If the sample contains substances with small differences in size, either decrease the sample volume, or decrease the flow rate (in very difficult cases, it may be necessary to decrease both).⁴ See [Adjusting pressure limits in chromatography system software, on page 5](#)⁵ pH range where resin can be operated without significant change in function.⁶ pH range where resin can be subjected to cleaning- or sanitization-in-place without significant change in function.

First time use

Connecting the column

| Step | Action |
|------|---|
| 1 | Before connecting the column to a chromatography system, start the pump to remove any air bubbles from the system, particularly in the tubing and valves. |
| 2 | Stop the pump. |
| 3 | Mount the column vertically, remove the stop plug and connect the inlet tubing to the system "drop-to-drop". |

Step Action

| | |
|---|---|
| 4 | Remove the transport device and connect the column outlet tubing to, for example, a monitor cell. Save the transport device for use when storing the column. The column is now ready for use. |
|---|---|

Equilibrating the column

Tip: *Equilibrate the column a day before usage to save time.*

Make sure that an appropriate pressure limit has been set. Equilibrate the column for first time use, or after long-term storage as follows:

Step Action

| | |
|---|---|
| 1 | One column volume (CV) of low ionic strength buffer at 1 mL/min for 16/600 or 2.6 mL/min for 26/600. |
| 2 | Two CV buffer, for example, 0.05 M sodium phosphate, 0.15 M NaCl, pH 7.2 at 1.6 mL/min for 16/600 or 4.3 mL/min for 26/600. |

Recommended running conditions

| | |
|--|---|
| Recommended operating flow rate ¹ | 1 mL/min for 16/600 or 2.6 mL/min for 26/600 |
| Sample volume | 0.5% to 4% of the CV (0.6 to 4.8 mL for 16/600 or 1.6 to 12.8 mL for 26/600) |
| Note: <i>Sample volume is critical for the separation.</i> | |
| Sample preparation | Dissolve the sample in running buffer, filter through 0.22 µm filter or centrifuge at 10 000 x g for 10 min |
| Buffer | 0.05 M NaPO ₄ , 0.15 M NaCl, pH 7.2 or select a buffer appropriate for the next purification step. To avoid pH dependent nonionic interactions with the matrix, include at least 0.15 M salt in the buffer (or use a buffer with equivalent ionic strength). |
| Regeneration | Regenerate the column after each run with one CV of running buffer at 1 mL/min for 16/600 or 2.6 mL/min for 26/600 |

¹ At room temperature in H₂O

Read [Optimizing, on page 3](#) for information on how to optimize a separation.

Note: *When running under cold conditions or using buffer with high viscosity, adjust the flow rate so that the back pressure limit is not exceeded.*

Delivery and storage

The prepacked column is delivered in 0.2 M sodium acetate, 20% ethanol (Superdex 30 and Superdex 75) or 20% ethanol (Superdex 200). If the column needs to be stored for more than two days after use, wash the column with four CV distilled water, and then equilibrate with four CV 0.2 M sodium acetate, 20% ethanol or 20% ethanol only, depending on the resin. Use the transport device to prevent air from entering the column and destroying the column packing. Connect the transport device to the capillary tubing at the column outlet. Start the pump and fill the device up to approximately 50% of the total device volume.



Daily use

Stable to commonly used aqueous buffers, pH 3 to 12



Cleaning

Acetonitrile, up to 30%

NaOH, up to 0.5 M

Ethanol, up to 70% (Superdex 30 pg)

Ethanol, up to 24% (Superdex 75 pg and Superdex 200 pg)

Acetic acid, up to 1 M

Isopropanol, up to 30%

Guanidine hydrochloride, up to 6 M


Urea, up to 8 M

Hydrochloric acid, up to 0.1 M (Superdex 30 pg)




Avoid

Unfiltered solutions



Column: HiLoad 16/600 Superdex 200 pg
 Sample: Solution of transferrin (M_r 81 000) and IgG (M_r 160 000) by equal weight.
 Sample concentration: 8 mg/mL
 Buffer: 50 mM NaPO₄, 0.1 M NaCl, pH 7.2
 Flow rate: 1 mL/min

Fig 4. Influence of sample volume on column resolution.



Column: HiLoad 16/600 Superdex 30 pg
 Sample: IGF-1 containing monomers and dimers
 Sample concentration: a) 1.25 mg/mL
 b) 5 mg/mL
 Sample volume: 1 mL (0.8% of CV)
 Buffer: 50 mM sodium acetate, 0.1 M NaCl, pH 5.0

Fig 5. Influence of flow rate on the column resolution.

Column resolution is calculated as:

$$R_s = \frac{2(V_{R2} - V_{R1})}{W_{b2} + W_{b1}}$$

where,

V_{R1} = Retention (elution) volume of the first peak

V_{R2} = Retention (elution) volume of the second peak

W_{b1} = Base width of the first peak

W_{b2} = Base width of the second peak

V_R and W_b in same units.




Fig 3. Dimensions of the column.

Optimizing

Perform a first run as described in [Recommended running conditions, on page 2](#). If the results obtained are unsatisfactory, consider the following:

| Action | Effect |
|------------------------|---------------------|
| Decrease flow rate | Improved resolution |
| Decrease sample volume | Improved resolution |

[Fig. 4, on page 3](#) and [Fig. 5, on page 3](#) demonstrate the influence of sample volume and flow rate on the resolution.

Cleaning-In-Place (CIP)

Regular cleaning

Wash the column with one-half to one CV 0.5 M NaOH at a flow rate of 0.8 mL/min for 16/600 or 2.2 mL/min for 26/600 to remove most of the nonspecifically bound proteins from the chromatography resin.

After cleaning, immediately equilibrate the column with at least two CV buffer. Further equilibration is necessary if the buffer contains detergents. Wait until the UV baseline stabilizes before starting a new purification.

More rigorous cleaning

Wash the column at a flow rate of 0.8 mL/min for 16/600 or 2.2 mL/min for 26/600 at room temperature with the following solutions:

| Step | Action |
|------|---|
| 1 | Four CV 0.5 M NaOH (removes hydrophobic proteins or lipoproteins) followed by four CV distilled water. |
| 2 | 2 One-half CV 30% isopropanol (removes lipids and very hydrophobic proteins), followed by two CV distilled water. |

Before starting a new purification, equilibrate the column after cleaning with at least five CV running buffer.

Changing the adapter net ring

After following the cleaning procedures above, if the back pressure of the column remains too high, change the net ring in the column adapter. Follow the instructions below thoroughly since column efficiency is easily impaired if handled without care. Use distilled water as a liquid. For an exploded view of the adapter, see [Fig. 8, on page 5](#).

| Step | Action |
|------|--|
| 1 | Close the outlet tubing of the column with a stop plug, and mark the level of the chromatography resin surface on the glass tube using a colored pen. |
| 2 | Slacken the adapter O-ring slightly by turning the black adjusting knob counter-clockwise. |
| | Note: <i>It should still seal against the glass wall but allow the adapter to slide. Unscrew the top piece from the column.</i> |
| 3 | Connect the adapter to the pump and start pumping at a flow rate of 1 mL for 16/600 or 2.6 mL/min for 26/600. Allow the flow to push the adapter upwards. |
| 4 | When the glass tube is completely full, take out the adapter and stop the pump. The glass tube should be completely filled with liquid. |
| 5 | Change the adapter net ring. |
| 6 | To avoid any air bubbles under the net, inject 20% ethanol through the adapter using a syringe. |
| 7 | Insert the adapter into the column at an angle of 45°, avoiding air bubbles. Slide the plunger 1 to 2 cm down and tighten the O-ring. Remove excess liquid completely before screwing the top piece onto the column end piece. |
| 8 | Remove the syringe and slide down the adapter until it touches the chromatography resin surface. Tighten the O-ring and reconnect the inlet tubing to the system, avoiding air bubbles. |

Step Action

| | |
|----|---|
| 9 | Remove the stop plug and start the pump. Increase the flow rate until the resin surface is approximately 3 mm above the pen mark. Stop the pump and close the outlet tubing with the stop plug again. |
| | Note: <i>This step requires a pump with high flow rate capacity up to a pressure of 0.5 MPa (5 bar).</i> |
| 10 | Disconnect the inlet tubing and slacken the adapter O-ring slightly by turning the adjusting knob counter-clockwise. Press the adapter downwards up to the pen mark. Tighten the O-ring. |
| | Note: <i>Do not loosen the O-ring too much as this will result in chromatography resin passing through the O-ring.</i> |
| 11 | Reconnect the inlet tubing and avoid introducing air into the system. |

Troubleshooting

| Symptom | Remedy |
|---|---|
| Increased back pressure over the column | Clean the column according to the section Cleaning-In-Place (CIP), on page 4 |
| Loss of resolution and/or decreased sample recovery | Clean the column according to the section Cleaning-In-Place (CIP), on page 4 |
| Air bubbles in the column | Reverse the direction of flow and pump five CV of degassed water through the column at the same flow rate that was used during the run. |
| Space between adapter and resin | Close the outlet tubing with the stop plug and then disconnect the inlet tubing. Slacken the O-ring slightly by turning the adjusting knob counter-clockwise and push or screw the adapter down until it touches the resin surface. Tighten the O-ring. To maintain an airtight system, reconnect the inlet tubing immediately. |

Testing the column efficiency

Cytiva packs columns to the highest standards and each column is thoroughly tested with respect to the number of theoretical plates per meter (N/m) ([Fig. 6, on page 4](#)).




Fig 6. Column efficiency test

Sample: 2% acetone in water

Sample volume: 200 μ L(16/600) and 500 μ L (26/600)

Eluent: Distilled water
 Flow velocity: 60 cm/h
 Flow rate: 2.0 mL/min (16/600)
 5.3 mL/min (26/600)
 Temperature: Room temperature (25°C)

Column efficiency is calculated using the equation:

$$N/m = 5.54 \times \left(\frac{V_R}{W_h} \right)^2 / L$$

where,

V_R = Peak retention (elution) volume

W_h = Peak width at half peak height

L = Bed height (meter)

V_R and W_h have the same units.

Adjusting pressure limits in chromatography system software

Pressure generated by the flow, through a column, affects the packed bed, and the column hardware, see [Fig. 7, on page 5](#).

Increased pressures might be generated when running/using one or a combination of the following conditions:

- High flow rates
- Buffers or sample with high viscosity
- Low temperature
- A flow restrictor

Note: Exceeding the pressure limits (see [Table 2, on page 2](#)) will damage the column.




Fig 7. Precolumn and post-column measurements.

ÄKTA™ avant and ÄKTA pure

The system will automatically handle all pressure limits, which facilitates an optimal functionality without any need of adjustments.

ÄKTAexplorer, ÄKTApurifier, ÄKTA FPLC and other systems with pressure sensor in the pump


To obtain optimal functionality, the pressure limits in the software may be adjusted according to the following procedure:

| Step | Action |
|------|--|
| 1 | Replace the column with a piece of tubing. Run the pump at the maximum intended flow rate. Note the pressure as <i>total system pressure</i> . |
| 2 | Disconnect the tubing and run the pump at the same flow rate used in step 1. Note that there will be a drip from the column valve. Note the pressure during this operation as <i>measured pressure</i> . |
| 3 | Calculate column pressure limit as a sum of total system pressure and Δp (pressure over the packed bed) (see Table 2, on page 2). |
| 4 | Replace the column pressure limit in the software with the calculated value. |

Calculate post-column pressure as the difference between total system pressure and measured pressure.

Column hardware pressure limit (see [Table 2, on page 2](#)) must never exceed the sum of post-column pressure and Δp .

Note: Repeat the procedure each time the parameters are changed.



| | |
|----|---|
| 1 | Net ring |
| 2 | Support screen |
| 3 | O-ring |
| 4 | Plunger |
| 5 | Ferrule |
| 6 | Capillary tubing |
| 7 | Inner shaft |
| 8 | Adapter shaft |
| 9 | Top end cap |
| 10 | Adjusting knob |
| 11 | HiTrap/HiPrep, 1/16" male connector for ÄKTA design |
| 12 | Stop plug |

Fig 8. Exploded view of the XK column adapter used at the top of the HiLoad column.

Intended use

HiLoad Superdex columns are intended for research use only, and shall not be used in any clinical or *in vitro* procedures for diagnostic purposes.

Ordering information

| Product | Pack size | Product Code. |
|---------------------------------------|------------|---------------|
| HiLoad 16/600 Superdex 30 prep grade | 1 x 120 mL | 28989331 |
| HiLoad 26/600 Superdex 30 prep grade | 1 x 320 mL | 28989332 |
| HiLoad 16/600 Superdex 75 prep grade | 1 x 120 mL | 28989333 |
| HiLoad 26/600 Superdex 75 prep grade | 1 x 320 mL | 28989334 |
| HiLoad 16/600 Superdex 200 prep grade | 1 x 120 mL | 28989335 |
| HiLoad 26/600 Superdex 200 prep grade | 1 x 320 mL | 28989336 |

| Accessories | No. supplied | Product Code. |
|--|--------------|---------------|
| Accessory kit XK 16 ¹ | 1 | 28989978 |
| Accessory kit XK 26 ¹ | 1 | 28989979 |
| Support screen XK 16 | 5 | 19065101 |
| Support screen XK 26 | 5 | 18937701 |
| Net ring (10 µm) XK 16 | 5 | 18876101 |
| Net ring (10 µm) XK 26 | 5 | 18876001 |
| O-ring XK 16 | 5 | 19016301 |
| O-ring XK 26 | 5 | 28978227 |
| Stop plug female, 1/16" | 5 | 11000464 |
| HiTrap/HiPrep 1/16" male connector for ÄKTA design | 8 | 28401081 |
| Transport device | 1 | 18117643 |

¹ Accessory kits XK 16 and XK 26 are suitable for repacking purposes and contain: 2 support screens, 5 net rings, 2 O-rings, 2 stop plugs, 10 HiTrap/HiPrep 1/16" male connectors for ÄKTA design, and 1 tool for dismantling.

| Related literature | Product Code. |
|--|---------------|
| Size Exclusion Chromatography: Principles and Methods, Handbook, | 18102218 |

cytiva.com/protein-purification

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate.

ÄKTA, HiLoad, and Superdex are trademarks of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva.

All other third-party trademarks are the property of their respective owners.

© 2020 Cytiva

All goods and services are sold subject to the terms and conditions of sale of the supplying company operating within the Cytiva business. A copy of those terms and conditions is available on request. Contact your local Cytiva representative for the most current information.

For local office contact information, visit cytiva.com/contact

28992017 AF V:9 10/2020

| Related literature | Product Code. |
|--|---------------|
| Size exclusion chromatography columns and resins, Selection Guide | 18112419 |
| Prepacked chromatography columns for ÄKTA systems, Selection Guide | 28931778 |

