

Spare parts recommended to keep on site for FlexStand™ Standard Holders FS-01S and FS- 01S-30 and FlexStand Standard Systems FS-02RLP and FS-02RLP/50

Seals and tubing are considered as consumables and should be replaced annually or when needed. Inspect the seals and tubing for wear regularly. Please keep a set of each in your stock to prevent long downtime. For code numbers refer to the table below.

Description	Quantity / pack	Material	Code No.	Model no
Gasket 50 mm i.d. 38 mm (1.5 inch)	4	Silicone	56-4109-96	KG12S
Tubing i.d. 6.4 mm (0.25 inch)	Length=3.1 m (10 feet)	Bioprene™	56-4106-18	PTPM06-10
Tubing i.d. 9.5 mm (0.375 inch)	Length=3.1 m (10 feet)	Bioprene	56-4106-19	PTPM09-10
Tubing i.d. 13 mm (0.5 inch)	Length =7.6 m (25 feet)	Bioprene	56-4106-22	PTPM12



imagination at work

GE, imagination at work, and GE monogram are trademarks of General Electric Company.
FlexStand is a trademark of GE Healthcare companies.
Bioprene is a trademark of Nam Liang Enterprise Co., Ltd
ÄKTA is a trademark of GE Healthcare companies.
© 2011 General Electric Company - All rights reserved.
GE Healthcare Bio-Sciences AB, Björkgatan 30, 751 84 Uppsala, Sweden.