Procedure

Preparation of Biacore regeneration buffers of different pH

Mix buffers according to indicated volumes (mL) to achieve target buffer pH.

Target buffer (pH)	Source buffer Glycine pH 1.5 (mL)	Source buffer Glycine pH 2.0 (mL)
1.6	7.0	3.0
1.7	4.6	5.4
1.8	2.7	7.3
1.9	1.2	8.8
	Glycine pH 2.0 (mL)	Glycine pH 2.5 (mL)
2.1	7.0	3.0
2.2	4.6	5.4
2.3	2.7	7.3
2.4	1.2	8.8
	Glycine pH 2.5 (mL)	Glycine pH 3.0 (mL)
2.6	7.0	3.0
2.7	4.6	5.4
2.8	2.7	7.3
2.9	1.2	8.8

Ordering information

Product name	Description	Contents	Product code
Glycine 1.5	10 mM glycine-HCl, pH 1.5	1 × 100 mL	BR100354
Glycine 2.0	10 mM glycine-HCl, pH 2.0	1 × 100 mL	BR100355
Glycine 2.5	10 mM glycine-HCl, pH 2.5	1 × 100 mL	BR100356
Glycine 3.0	10 mM glycine-HCl, pH 3.0	1 × 100 mL	BR100357

Note: solutions should be stored at +4°C to +8°C

cytiva.com

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate. Biacore is a trademark of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva.

© 2020 Cytiva

All goods and services are sold subject to the terms and conditions of sale of the supplying company operating within the Cytiva business. A copy of those terms and conditions is available on request. Contact your local Cytiva representative for the most current information.

For local office contact information, visit cytiva.com/contact

CY17750-08Dec20-PD

cytiva