

# Preventive maintenance

Ensure optimal performance of your Palltronic Flowstar IV and Palltronic Flowstar V filter integrity test instruments

Preventive maintenance (PM) reduces the risk of unexpected breakdowns, helps keep your schedule on track, and ensures your equipment remains properly calibrated. PM is built into our equipment design process for long-term reliability and backed by extensive global experience. Our certified service engineers are trained to deliver expert PM tailored to your system. On completion, we provide detailed service reports for full traceability.

## Comprehensive system assessment

During PM, our qualified engineers perform comprehensive tests and verifications to ensure your Palltronic Flowstar IV and Palltronic Flowstar V integrity test instruments meet original factory specifications. They replace any necessary components using quality parts from Cytiva. These procedures are based on protocols optimized during system design and reflect the same functional and quality checks applied during assembly and pre-delivery validation.

### Tests and inspections

---

Pump flow rate test

---

Pressure sensor test

---

Temperature sensor test

---

Valve test

---

Software check, including software updates for non-validated instruments

---

Load cell test

---

Mixer test

---

Mechanical inspection

---

User interface functional test

---

### Parts replaced

---

Feed pump

---

Peristaltic tubes

---

Exchange of wear-and-tear kit

---

Valves

---

Tank air filter

---

Sanitary clamp gaskets (AKTA flux™ 6 system)

---

Valve diaphragms

---

Tygon tube kit (AKTA flux s system)

---

Pressure sensor diaphragms

---

Optional pumps

---

Mixer stirrer

---



Fig 1. Palltronic Flowstar V integrity test instrument.

## Documentation

We provide a standardized set of documentation and test results after each service or repair. This ensures full traceability, saves time, reduces costs, and supports compliance in regulated environments. Our service engineers document every PM visit and issue a service report upon completion.

## Frequency

System usage	Type of PM visit
Normal usage: up to 2000 hours per year	One PM visit every second year
High usage: 2000 to 4000 hours per year, or use with aggressive solvents	One PM visit annually
Continuous usage: more than 4000 hours per year	Two PM visits annually

## What we do and why it matters

Main components	Outcomes
Functional testing, inspections, and calibration	Ensures reliable accuracy of testing data and system functionality.
Wear-and-tear replacement	Minimizes unexpected failure risk. Extends the life of your equipment.
Exchange of all elastomers in contact with process fluid	Minimizes contamination risk due to leakage to ensure high quality end product for human use.
As-found/as-left testing	Ensures equipment runs according to specifications after completed PM.
Documentation	Signed PM service report upon completion.

## An integral part of our service plans

To maintain consistent system performance of your Palltronic Flowstar IV and Palltronic Flowstar V integrity test instruments, PM is included in all our service agreements.



Extended  
life cycle



Service  
traceability



Maximized  
productivity

Learn more about how our service agreements can support your operations at [cytiva.com/equipment-services](https://cytiva.com/equipment-services)

## cytiva.com

Cytiva and the Drop logo are trademarks of Life Sciences IP Holdings Corporation or an affiliate doing business as Cytiva. ÄKTA and ÄKTA flux are trademarks of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva. Any other trademarks are the property of their respective owners.

© 2025 Cytiva

For local office contact information, visit [cytiva.com/contact](https://cytiva.com/contact)

CY35819-28Nov25-FL

